

Certificate of Accuracy

Cal	ihration	Certificate	No	218233
	IDIGUUII	CELLICALE	INU-	

Radar Details:

AGD 340-505-000 Serial No. 117389-0025

Date of Calibration: 28 December 2023

Marsh Metrology Calibration

Procedure: MC-672

Statement:

Marsh Metrology certifies that the speed measurement radar detailed above was calibrated and tested on the date shown; that it is functioning properly and that it can accurately measure the speed of a moving motor vehicle within +/-1 km/hr.

Authorised signatory:

Signature: Date: 28 December 2023

Name: Ron Bake

Position: General Manager