

Certificate of Accuracy

Cali	hration	Certificate	No	218462
Lan	vialivii	CELLICALE	INU-	ムエいてい と

Radar Details:

AGD 340-505-000 Serial No. 117387-0021

Date of Calibration: 08 January 2024

Marsh Metrology Calibration

Procedure: MC-672

Statement:

Marsh Metrology certifies that the speed measurement radar detailed above was calibrated and tested on the date shown; that it is functioning properly and that it can accurately measure the speed of a moving motor vehicle within +/-1 km/hr.

Authorised signatory:

Signature: Softe Date: 08 January 2024

Name: Ron Bake

Position: General Manager