

Certificate of Accuracy

| Calibration Certificate No. | 220104 |
|-----------------------------|--------|
|-----------------------------|--------|

Radar Details:

AGD 340-505-000 Serial No. 122596-0006

Date of Calibration: 28 February 2024

Marsh Metrology Calibration

Procedure: MC-672

Statement:

Marsh Metrology certifies that the speed measurement radar detailed above was calibrated and tested on the date shown; that it is functioning properly and that it can accurately measure the speed of a moving motor vehicle within +/-1 km/hr.

Authorised signatory:

Signature: Softe Date: 28 February 2024

Name: Ron Bake

Position: General Manager