

Certificate of Accuracy

Calibration	Certificate No.	231896
Calibi atibii	CELLITICALE NO.	ZJ1030

Radar Details:

AGD 340-505-000 Serial No. 126179-0008

Date of Calibration: 30 December 2024

Marsh Metrology Calibration

Procedure: MC-672

Statement:

Marsh Metrology certifies that the speed measurement radar detailed above was calibrated and tested on the date shown; that it is functioning properly and that it can accurately measure the speed of a moving motor vehicle within +/- 1 km/hr.

Authorised signatory:

Signature: Date: 30 December 2024

Name: Ron Bake

Position: General Manager